NOTICE: This document contains references to Varian. Please note that Varian, Inc. is now part of Agilent Technologies. For more information, go to **www.agilent.com/chem.**

Agilent Technologies

Varian, Inc.
2700 Mitchell Drive
Walnut Creek, CA 94598-1675/USA

Perkin Elmer GC drivers

Release Notes

Release Notes

HS-40

The ‘time events capability’ option is not available.

The MHE mode requires adding the 800 MIB relay control driver in the system configuration. The 800 MIB relay should be connected to the Ready In of the HS-40.

In MHE mode, the relay program should be the following:

Prerun state: OPEN

Injection: CLOSE

At 0.00 min: OPEN

In Normal mode, use the following relay program:

Prerun state: CLOSE

Injection: CLOSE

At 0.00 min: CLOSE